

3160007282 Questions and Answer

IFB # 3160007282 Customized Testing and Reporting Services Human DNA Testing for MSFL

April 8, 2025

Question 1:

Will MSFL accept Promega's Maxwell 16 for bone extractions?

Answer: Yes

Question 2:

Will MSFL accept Promega's Maxwell RSC 48 for evidence extractions?

Answer: No

Question 3:

Is the ability to perform mitochondrial DNA testing a requirement for the project?

Answer: No

Question 4:

Can FIGG/SNP testing, database upload and records searching be outsourced to another lab in the US?

Answer: No

Question 5:

Will MSFL accept data using Yfiler PCR amplification kit instead of Yfiler Plus PCR kit?

Answer: No

I acknowledge and accept all terms and conditions of the Invitation for Bid (IFB), Amendments, and Question & Answer. I hereby, certify that I am authorized to sign for my company.

Signature

Date

Name (Printed)

Date